

| Nominal <i>(Nominal)</i> | Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 4 °C                     | 3.900 °C                 | 4.0 °C           | 0.100 °C             | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 8 °C                     | 8.022 °C                 | 7.7 °C           | -0.322 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 30 °C                    | 30.042 °C                | 29.6 °C          | -0.442 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 50 %HR                   | 50.621 %HR               | 53.2 %HR         | 2.579 %HR            | ± 5.0 %HR                   | ± 1.9E+00 %HR                      |