

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	0.030 °C	0.3 °C	0.270 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	19.955 °C	20.2 °C	0.245 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.042 °C	30.2 °C	0.158 °C	± 0.5 °C	± 4.0E-01 °C
50 °C	50.007 °C	50.0 °C	-0.007 °C	± 5.0 °C	± 1.1E+00 °C