

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.072 °C	29.9 °C	-0.172 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	28.5 %HR	-2.070 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.826 %HR	49.1 %HR	-1.726 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	67.9 %HR	-2.548 %HR	± 3.0 %HR	± 1.9E+00 %HR