

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.102 °C	29.9 °C	-0.202 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	29.9 %HR	-0.670 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	50.202 %HR	52.7 %HR	2.498 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	70.570 %HR	71.5 %HR	0.930 %HR	± 2.5 %HR	± 1.9E+00 %HR