

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.6	0.59873	1.27	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.69873	1.27	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.79873	1.27	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.89872	1.28	± 1.9	± 4.3E-01 mm					
0 a 1	0.99872	1.28	± 2	± 4.3E-01 mm					
0 a 1.1	1.09872	1.28	± 2	± 4.8E-01 mm					
0 a 1.2	1.19871	1.29	± 2.1	± 4.8E-01 mm					
0 a 1.3	1.29871	1.29	± 2.2	± 4.8E-01 mm					
0 a 1.4	1.39870	1.30	± 2.3	± 4.8E-01 mm					
0 a 1.5	1.49870	1.30	± 2.4	± 4.8E-01 mm					
0 a 1.6	1.59869	1.31	± 2.4	± 4.8E-01 mm					
0 a 1.7	1.69869	1.31	± 2.5	± 4.8E-01 mm					
0 a 1.8	1.79868	1.32	± 2.6	± 4.8E-01 mm					
0 a 1.9	1.899	1.33	± 2.7	± 4.8E-01 mm					
0 a 2	1.999	1.33	± 2.8	± 4.8E-01 mm					
0 a 2.1	2.099	1.34	± 2.8	± 4.8E-01 mm					