

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 0.6	0.599	1.27	± 1.6	± 4.3E-01 mm					
0 a 0.7	0.699	1.27	± 1.7	± 4.3E-01 mm					
0 a 0.8	0.799	1.27	± 1.8	± 4.3E-01 mm					
0 a 0.9	0.899	1.28	± 1.9	± 4.3E-01 mm					
0 a 1	0.999	1.28	± 2	± 4.3E-01 mm					
0 a 1.1	1.099	1.28	± 2	± 4.8E-01 mm					
0 a 1.2	1.199	1.29	± 2.1	± 4.8E-01 mm					
0 a 1.3	1.299	1.29	± 2.2	± 4.8E-01 mm					
0 a 1.4	1.399	1.30	± 2.3	± 4.8E-01 mm					
0 a 1.5	1.499	1.30	± 2.4	± 4.8E-01 mm					
0 a 1.6	1.599	1.31	± 2.4	± 4.8E-01 mm					
0 a 1.7	1.699	1.31	± 2.5	± 4.8E-01 mm					
0 a 1.8	1.799	1.32	± 2.6	± 4.8E-01 mm					
0 a 1.9	1.899	1.00	± 2.7	± 4.8E-01 mm					
0 a 2	1.999	1.00	± 2.8	± 4.8E-01 mm					