

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.5	0.49958	0.420	± 0.8	±	4.3E-01 mm						
0 a 1	0.99961	0.390	± 1	±	4.3E-01 mm						
0 a 1.5	1.49947	0.530	± 1.2	±	4.8E-01 mm						
0 a 2	1.99955	0.450	± 1.4	±	4.8E-01 mm						
0 a 2.5	2.49948	0.520	± 1.6	±	4.8E-01 mm						
0 a 3	2.99939	0.610	± 1.8	±	4.8E-01 mm						
0 a 3.5	3.49950	0.500	± 2	±	4.8E-01 mm						
0 a 4	3.99928	0.720	± 2.2	±	4.8E-01 mm						
0 a 4.5	4.49918	0.820	± 2.4	±	4.8E-01 mm						
0 a 5	4.99899	1.010	± 2.6	±	4.8E-01 mm						