

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.3 μm	23.5 μm	0.2 μm	± 2.5 μm	±	5.8E-02 μm
51.6 μm	51.2 μm	-0.4 μm	± 2.5 μm	±	5.8E-02 μm
124.6 μm	123 μm	-1.6 μm	± 3.7 μm	±	6.1E-02 μm
250.3 μm	251 μm	0.7 μm	± 7.5 μm	±	6.9E-02 μm
486.0 μm	485 μm	-1.0 μm	± 15 μm	±	9.4E-02 μm
963.0 μm	964 μm	1.0 μm	± 29 μm	±	1.6E-01 μm