

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 1 mm              | 0.98 mm   | -20 µm        | ± 50 µm           | ±                           | 5.8E+00 µm |
| 2 mm              | 1.98 mm   | -20 µm        | ± 50 µm           | ±                           | 5.8E+00 µm |
| 3 mm              | 2.98 mm   | -20 µm        | ± 50 µm           | ±                           | 5.8E+00 µm |
| 5 mm              | 4.99 mm   | -10 µm        | ± 50 µm           | ±                           | 5.8E+00 µm |
| 10 mm             | 9.99 mm   | -10 µm        | ± 50 µm           | ±                           | 5.8E+00 µm |
| 15 mm             | 15.00 mm  | 0 µm          | ± 50 µm           | ±                           | 5.8E+00 µm |
| 20 mm             | 20.00 mm  | 0 µm          | ± 50 µm           | ±                           | 5.8E+00 µm |
| 25 mm             | 25.00 mm  | 0 µm          | ± 50 µm           | ±                           | 5.8E+00 µm |