

| Tipo de Medida (Measurement Type) | Patrón (Standard) | UBP (UT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |
|-----------------------------------|-------------------|----------|---------------|-------------------|-----------------------------|
| Profundidad / (Depth)             | 1 mm              | 0.99 mm  | -10 µm        | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 2 mm              | 1.50 mm  | 0 µm          | ± 30 µm           | ± 5.8E+00 µm                |
| Profundidad / (Depth)             | 2 mm              | 1.99 mm  | -10 µm        | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 3 mm              | 3.00 mm  | 0 µm          | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 5 mm              | 4.99 mm  | -10 µm        | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 10 mm             | 9.99 mm  | -10 µm        | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 15 mm             | 15.00 mm | 0 µm          | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 20 mm             | 19.99 mm | -10 µm        | ± 30 µm           | ± 6.1E+00 µm                |
| Profundidad / (Depth)             | 25 mm             | 24.99 mm | -10 µm        | ± 30 µm           | ± 6.1E+00 µm                |