

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Profundidad / (Depth)	1 mm	1.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	2 mm	1.52 mm	20 μm	± 30 μm	± 5.8E+00 μm
Profundidad / (Depth)	2 mm	2.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	3 mm	3.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	5 mm	5.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	10 mm	10.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	15 mm	15.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	20 mm	20.02 mm	20 μm	± 30 μm	± 6.1E+00 μm
Profundidad / (Depth)	25 mm	25.02 mm	20 μm	± 30 μm	± 6.1E+00 μm