

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.002 °C	30.2 °C	0.198 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.550 %HR	33.5 %HR	2.950 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	52.513 %HR	54.3 %HR	1.787 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.441 %HR	69.5 %HR	-2.941 %HR	± 3.0 %HR	± 1.9E+00 %HR