

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.002 °C	30.0 °C	-0.002 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.550 %HR	33.2 %HR	2.650 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	52.513 %HR	54.8 %HR	2.287 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.441 %HR	70.7 %HR	-1.741 %HR	± 3.0 %HR	± 1.9E+00 %HR