

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.002 °C	30.0 °C	-0.002 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.550 %HR	33.5 %HR	2.950 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	52.513 %HR	55.2 %HR	2.687 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	72.441 %HR	70.3 %HR	-2.141 %HR	± 3.0 %HR	± 1.9E+00 %HR