

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.631 °C	-25.3 °C	0.331 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.220 °C	0.5 °C	0.280 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.104 °C	5.2 °C	0.096 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.3 °C	0.162 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.1 °C	0.092 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	29.772 °C	29.9 °C	0.128 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	39.659 °C	39.8 °C	0.141 °C	± 0.45 °C	± 4.0E-01 °C

