

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.631 °C	-25.3 °C	0.331 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.220 °C	0.5 °C	0.280 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.104 °C	5.4 °C	0.296 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.3 °C	0.162 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.1 °C	0.092 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	29.772 °C	29.8 °C	0.028 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	39.659 °C	39.4 °C	-0.259 °C	± 0.45 °C	± 4.0E-01 °C

