

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.631 °C	-25.2 °C	0.431 °C	± 0.95 °C	± 4.0E-01 °C
0 °C	0.220 °C	0.0 °C	-0.220 °C	± 0.45 °C	± 4.0E-01 °C
5 °C	5.104 °C	5.2 °C	0.096 °C	± 0.45 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.1 °C	-0.038 °C	± 0.45 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.2 °C	0.192 °C	± 0.45 °C	± 4.0E-01 °C
30 °C	29.772 °C	29.7 °C	-0.072 °C	± 0.45 °C	± 4.0E-01 °C
40 °C	39.659 °C	39.9 °C	0.241 °C	± 0.45 °C	± 4.0E-01 °C

