

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	31.787 °C	31.7 °C	-0.087 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.038 %HR	37.6 %HR	2.562 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.486 %HR	51.2 %HR	1.714 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.378 %HR	69.6 %HR	-1.778 %HR	± 3.0 %HR	± 1.9E+00 %HR