

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	31.787 °C	31.8 °C	0.013 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.038 %HR	37.9 %HR	2.862 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.486 %HR	51.4 %HR	1.914 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.378 %HR	70.2 %HR	-1.178 %HR	± 3.0 %HR	± 1.9E+00 %HR