

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.368 °C	10.4 °C	0.032 °C	± 0.12 °C	± 4.0E-01 °C
20 °C	19.814 °C	19.9 °C	0.086 °C	± 0.12 °C	± 4.0E-01 °C
30 °C	29.731 °C	29.8 °C	0.069 °C	± 0.12 °C	± 4.0E-01 °C
40 °C	39.759 °C	39.8 °C	0.041 °C	± 0.12 °C	± 4.0E-01 °C
-	-	-	-	-	-