

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.897 °C	10.1 °C	0.203 °C	± 0.23 °C	± 4.0E-01 °C
20 °C	19.804 °C	20.0 °C	0.196 °C	± 0.23 °C	± 4.0E-01 °C
30 °C	29.852 °C	30.0 °C	0.148 °C	± 0.23 °C	± 4.0E-01 °C
40 °C	39.899 °C	40.0 °C	0.101 °C	± 0.23 °C	± 4.0E-01 °C
-	-	-	-	-	-