

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	9.766 °C	9.8 °C	0.034 °C	± 0.12 °C	± 4.0E-01 °C
20 °C	19.814 °C	19.9 °C	0.086 °C	± 0.12 °C	± 4.0E-01 °C
30 °C	29.832 °C	29.9 °C	0.068 °C	± 0.12 °C	± 4.0E-01 °C
40 °C	39.909 °C	40.0 °C	0.091 °C	± 0.12 °C	± 4.0E-01 °C