

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.852 °C	29.9 °C	0.048 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.323 %HR	30.5 %HR	1.177 %HR	± 2.0 %HR	± 1.9E+00 %HR
50 %HR	49.118 %HR	49.6 %HR	0.482 %HR	± 2.0 %HR	± 1.9E+00 %HR
70 %HR	73.433 %HR	75.4 %HR	1.967 %HR	± 2.0 %HR	± 1.9E+00 %HR
-	-	-	-	-	-