

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
18 °C	20.155 °C	20.1 °C	-0.055 °C	± 0.5 °C	± 4.0E-01 °C
23 °C	23.043 °C	23.1 °C	0.057 °C	± 0.5 °C	± 4.0E-01 °C
28 °C	30.042 °C	29.8 °C	-0.242 °C	± 0.5 °C	± 4.0E-01 °C
45 %HR	45.345 %HR	46.9 %HR	1.555 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	61.041 %HR	62.4 %HR	1.359 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.947 %HR	73.2 %HR	-1.747 %HR	± 3.0 %HR	± 1.9E+00 %HR