

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.631 °C	-25.4 °C	0.231 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.331 °C	-0.2 °C	0.131 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.104 °C	5.3 °C	0.196 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.3 °C	0.162 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.3 °C	0.292 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	31.556 °C	31.8 °C	0.244 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.659 °C	39.9 °C	0.241 °C	± 0.5 °C	± 4.0E-01 °C

