

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.631 °C	-25.3 °C	0.331 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.331 °C	0.0 °C	0.331 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.104 °C	5.1 °C	-0.004 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.1 °C	-0.038 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.0 °C	-0.008 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	31.556 °C	31.7 °C	0.144 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.659 °C	39.9 °C	0.241 °C	± 0.5 °C	± 4.0E-01 °C

