

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.631 °C	-26.0 °C	-0.369 °C	± 0.70 °C	± 4.0E-01 °C
0 °C	-0.331 °C	-0.6 °C	-0.269 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.104 °C	4.8 °C	-0.304 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.1 °C	-0.038 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	25.1 °C	0.092 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	31.556 °C	31.8 °C	0.244 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.659 °C	39.9 °C	0.241 °C	± 0.5 °C	± 4.0E-01 °C

