

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.631 °C	-25.0 °C	0.631 °C	± 1.1 °C	± 4.0E-01 °C
0 °C	0.220 °C	-0.1 °C	-0.320 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	5.104 °C	5.0 °C	-0.104 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.138 °C	10.0 °C	-0.138 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	24.9 °C	-0.108 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.772 °C	30.0 °C	0.228 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.659 °C	39.9 °C	0.241 °C	± 0.5 °C	± 4.0E-01 °C

