

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
50.60 μm	50.60 μm	0.00 μm	± 2.5 μm	±	5.8E-01 μm
144.40 μm	144.2 μm	-0.20 μm	± 2.5 μm	±	5.8E-01 μm
245.20 μm	245.8 μm	0.60 μm	± 2.5 μm	±	5.8E-01 μm
503.10 μm	503.0 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm