

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.005 mm	1.006 mm	1 μm	± 50.25 μm	±	5.8E-01 μm
1.5 mm	1.500 mm	0 μm	± 75. μm	±	5.8E-01 μm
2 mm	2.000 mm	0 μm	± 100 μm	±	5.8E-01 μm
4 mm	3.998 mm	-2 μm	± 200 μm	±	5.8E-01 μm
5 mm	4.998 mm	-2 μm	± 250 μm	±	5.8E-01 μm