

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.102 °C	30.2 °C	0.098 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	29.2 %HR	-1.370 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	50.202 %HR	49.2 %HR	-1.002 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	70.570 %HR	72.2 %HR	1.630 %HR	± 2.5 %HR	± 1.9E+00 %HR