

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	31.787 °C	31.9 °C	0.113 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	35.038 %HR	38 %HR	2.962 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	49.486 %HR	53 %HR	3.514 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	71.378 %HR	74 %HR	2.622 %HR	± 5 %HR	± 1.9E+00 %HR