

Certificado No. (Certificate #): 52324
Fecha de Calibración (Calibration Date): 2024-12-18

Equipo (Instrument): Medidor de Espesores
Marca (Brand): DEFELSKO

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
3.919 mm	3.98 mm	0.06 mm	± μm	± 5.8E+00 μm
5.944 mm	5.95 mm	0.01 mm	± μm	± 5.8E+00 μm
10.934 mm	10.92 mm	-0.01 mm	± μm	± 5.8E+00 μm
15.935 mm	15.84 mm	-0.09 mm	± μm	± 5.8E+00 μm
17.959 mm	17.90 mm	-0.06 mm	± μm	± 5.8E+00 μm
18.931 mm	18.84 mm	-0.09 mm	± μm	± 5.8E+00 μm
20.935 mm	20.82 mm	-0.11 mm	± μm	± 5.8E+00 μm
23.930 mm	23.84 mm	-0.09 mm	± μm	± 5.8E+00 μm
49.937 mm	49.72 mm	-0.22 mm	± μm	± 5.8E+00 μm