

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.052 °C	29.0 °C	-1.052 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.619 %HR	31 %HR	-3.619 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	50.437 %HR	49 %HR	-1.437 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	71.225 %HR	73 %HR	1.775 %HR	± 5 %HR	± 1.9E+00 %HR