

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.052 °C	30.1 °C	0.048 °C	± 1.0 °C	± 4.0E-01 °C
35 %HR	34.619 %HR	35 %HR	0.381 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	50.437 %HR	51 %HR	0.563 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	71.225 %HR	72 %HR	0.775 %HR	± 5 %HR	± 1.9E+00 %HR