

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.052 °C	30.1 °C	0.048 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	34.619 %HR	37.2 %HR	2.581 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.437 %HR	53.1 %HR	2.663 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	71.225 %HR	68.7 %HR	-2.525 %HR	± 3.0 %HR	± 1.9E+00 %HR