

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
2 °C	2.567 °C	2.4 °C	-0.167 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	8.513 °C	8.7 °C	0.187 °C	± 0.5 °C	± 4.0E-01 °C
70 %HR	70.448 %HR	71.8 %HR	1.352 %HR	± 3.0 %HR	± 1.9E+00 %HR