

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.102 °C	30.3 °C	0.198 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	29.6 %HR	-0.970 %HR	± 2.5 %HR	± 1.9E+00 %HR
50 %HR	50.202 %HR	36.3 %HR	-13.902 %HR	± 2.5 %HR	± 1.9E+00 %HR
70 %HR	70.570 %HR	47.6 %HR	-22.970 %HR	± 2.5 %HR	± 1.9E+00 %HR