

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
125.00	mm	125.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
126.00	mm	126.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
127.00	mm	127.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
128.00	mm	128.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
130.00	mm	130.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
131.00	mm	131.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
133.00	mm	133.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
135.00	mm	135.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
140.00	mm	140.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
145.00	mm	145.000	mm	0 μm	± 6 μm	±	5.8E+00 μm
150.00	mm	150.000	mm	0 μm	± 6 μm	±	5.8E+00 μm