

Patrón <i>(Standard)</i>		UBP <i>(UUT)</i>		Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.00	mm	99.940	mm	-60 μm	± 6 μm	±	5.8E+00 μm
101.00	mm	100.940	mm	-60 μm	± 6 μm	±	5.8E+00 μm
102.00	mm	101.940	mm	-60 μm	± 6 μm	±	5.8E+00 μm
103.00	mm	102.940	mm	-60 μm	± 6 μm	±	5.8E+00 μm
105.00	mm	104.940	mm	-60 μm	± 6 μm	±	5.8E+00 μm
106.00	mm	105.940	mm	-60 μm	± 6 μm	±	5.8E+00 μm
108.00	mm	107.950	mm	-50 μm	± 6 μm	±	5.8E+00 μm
110.00	mm	109.950	mm	-50 μm	± 6 μm	±	5.8E+00 μm
115.00	mm	114.950	mm	-50 μm	± 6 μm	±	5.8E+00 μm
120.00	mm	119.950	mm	-50 μm	± 6 μm	±	5.8E+00 μm
125.00	mm	124.950	mm	-50 μm	± 6 μm	±	5.8E+00 μm