
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1201.1 RPM	1200 RPM	-1.1 RPM	± 1205 RPM	±	1.1E-01 RPM
1601.5 RPM	1600 RPM	-1.5 RPM	± 1600.1 RPM	±	1.1E-01 RPM
2001.06 RPM	2000 RPM	-1.1 RPM	± 2005.1 RPM	±	1.1E-01 RPM
2401.1 RPM	2400 RPM	-1.1 RPM	± 2401.1 RPM	±	1.1E-01 RPM
3001.8 RPM	3000 RPM	-1.8 RPM	± 3010.1 RPM	±	1.1E-01 RPM