

Patrón <i>(Standard)</i>	UBP <i>(UIT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
5498.76 RPM	5500 RPM	1.24 RPM	± 5501.1 RPM	±	7.8E-01 RPM