

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
25 °C	24.848 °C	24.7 °C	-0.148 °C	± 0.5 °C	± 4.0E-01 °C
# 25 °C	24.848 °C	24.6 °C	-0.248 °C	± 0.5 °C	± 7.4E-02 °C
# 50 °C	49.917 °C	49.5 °C	-0.417 °C	± 0.5 °C	± 7.4E-02 °C
# 70 °C	69.972 °C	69.7 °C	-0.272 °C	± 0.5 °C	± 7.4E-02 °C
25 %HR	23.413 %HR	22.8 %HR	-0.613 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	48.975 %HR	48.1 %HR	-0.875 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	69.425 %HR	68.2 %HR	-1.225 %HR	± 3 %HR	± 1.9E+00 %HR