

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.30 μm	23.20 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm
51.60 μm	51.5 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm
124.60 μm	124.5 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm
250.30 μm	250.2 μm	-0.10 μm	± 2.5 μm	±	5.8E-01 μm
486.00 μm	486.1000 μm	0.1000 μm	± 2.5 μm	±	5.8E-01 μm
963.00 μm	962.9000 μm	-0.1000 μm	± 2.5 μm	±	6.0E-01 μm