

Certificado No. (Certificate #): 52313

Equipo (Instrument):

Reloj Comparador

Fecha de Calibración (CalibrationDate): 2024-12-12

Marca (Brand):

Humbolt

| Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) | |
|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|------------|
| 0.0500 in | 0.0508 in | 0.0007 in | ± 0.002 in | ± | 1.8E-03 in |
| 0.1010 in | 0.1033 in | 0.0023 in | ± 0.002 in | ± | 4.1E-03 in |
| 0.1500 in | 0.1503 in | 0.0003 in | ± 0.002 in | ± | 1.8E-03 in |
| 0.2000 in | 0.2003 in | 0.0002 in | ± 0.002 in | ± | 1.8E-03 in |
| 0.3000 in | 0.2995 in | -0.0005 in | ± 0.002 in | ± | 2.4E-03 in |
| 0.5000 in | 0.4988 in | -0.0013 in | ± 0.002 in | ± | 3.0E-03 in |
| 1.0000 in | 0.9985 in | -0.0015 in | ± 0.002 in | ± | 2.5E-03 in |
| 1.5000 in | 1.4993 in | -0.0008 in | ± 0.002 in | ± | 2.1E-03 in |
| 1.9000 in | 1.9005 in | 0.0005 in | ± 0.002 in | ± | 1.5E-03 in |