

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.0500 in	0.0508 in	0.0007 in	± 0.002 in	±	1.8E-03 in
0.1010 in	0.1033 in	0.0023 in	± 0.002 in	±	4.1E-03 in
0.1500 in	0.1503 in	0.0003 in	± 0.002 in	±	1.8E-03 in
0.2000 in	0.2003 in	0.0002 in	± 0.002 in	±	1.8E-03 in
0.3000 in	0.2995 in	-0.0005 in	± 0.002 in	±	2.4E-03 in
0.5000 in	0.4988 in	-0.0013 in	± 0.002 in	±	3.0E-03 in
1.0000 in	0.9985 in	-0.0015 in	± 0.002 in	±	2.5E-03 in
1.5000 in	1.4993 in	-0.0008 in	± 0.002 in	±	2.1E-03 in
1.9000 in	1.9005 in	0.0005 in	± 0.002 in	±	1.5E-03 in