

Clase de Exactitud (Accuracy Class): Clase I (Class I)

---

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 0.04	0.04097	-0.012	± 0.016	± 3.5E-01 in					
0 a 4.08	4.08095	-0.011	± 1	± 3.5E-01 in					
0 a 6.03	6.03118	-0.014	± 1.4	± 3.5E-01 in					
0 a 10.06	10.06083	-0.010	± 2.2	± 3.5E-01 in					
0 a 14.02	14.02097	-0.012	± 3	± 3.5E-01 in					
0 a 19.04	19.04072	-0.009	± 4	± 3.5E-01 in					
0 a 23.05	23.05077	-0.009	± 4.8	± 3.5E-01 in					
0 a 27.08	27.08065	-0.008	± 5.6	± 3.5E-01 in					
0 a 31.1	31.10062	-0.007	± 6.4	± 3.5E-01 in					
0 a 35.12	35.12080	-0.010	± 7.2	± 3.5E-01 in					
0 a 37.09	37.09076	-0.009	± 7.6	± 3.5E-01 in					
0 a 40.07	40.07085	-0.010	± 8.2	± 3.5E-01 in					
0 a 45.11	45.11089	-0.011	± 9.2	± 3.5E-01 in					
0 a 47.14	47.14085	-0.010	± 9.6	± 3.5E-01 in					
0 a 50.05	50.05115	-0.014	± 10.2	± 3.5E-01 in					

---