

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 0.04	0.04087	-0.010	± 0.016	± 3.5E-01 in					
0 a 4.08	4.08085	-0.010	± 1	± 3.5E-01 in					
0 a 6.03	6.03088	-0.011	± 1.4	± 3.5E-01 in					
0 a 10.06	10.06093	-0.011	± 2.2	± 3.5E-01 in					
0 a 14.02	14.02097	-0.012	± 3	± 3.5E-01 in					
0 a 19.06	19.06072	-0.009	± 4	± 3.5E-01 in					
0 a 23.1	23.10087	-0.010	± 4.8	± 3.5E-01 in					
0 a 27.05	27.05085	-0.010	± 5.6	± 3.5E-01 in					
0 a 31.1	31.10062	-0.007	± 6.4	± 3.5E-01 in					