

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
ft	ft	in	in	(Uncertainty)	ft	ft	in	in	(Uncertainty)
0 a 0.04	0.04097	-0.012	± 0.016	± 2.9E-01 in					
0 a 1.2	1.20095	-0.011	± 0.4	± 2.9E-01 in					
0 a 3.5	3.50118	-0.014	± 0.9	± 2.9E-01 in					
0 a 6.49	6.49082	-0.010	± 1.4	± 2.9E-01 in					
0 a 9.01	9.01096	-0.012	± 2	± 2.9E-01 in					
0 a 12.51	12.51070	-0.008	± 2.7	± 2.9E-01 in					
0 a 15.49	15.49075	-0.009	± 3.2	± 2.9E-01 in					
0 a 18.52	18.52063	-0.008	± 3.9	± 2.9E-01 in					
0 a 22.53	22.53059	-0.007	± 4.7	± 2.9E-01 in					
0 a 26.12	26.12077	-0.009	± 5.4	± 2.9E-01 in					
0 a 30.03	30.03073	-0.009	± 6.2	± 2.9E-01 in					
0 a 34.1	34.10083	-0.010	± 7	± 2.9E-01 in					
0 a 38.1	38.10086	-0.010	± 7.8	± 2.9E-01 in					
0 a 42.05	42.05084	-0.010	± 8.6	± 2.9E-01 in					
0 a 46.06	46.06114	-0.014	± 9.4	± 2.9E-01 in					
0 a 50.2	50.20095	-0.011	± 10.2	± 2.9E-01 in					
0 a 54.24	54.24125	-0.015	± 11	± 2.9E-01 in					
0 a 58.35	58.35037	-0.004	± 11.8	± 2.9E-01 in					
0 a 62.27	62.27024	-0.003	± 12.6	± 2.9E-01 in					
0 a 66.58	66.58085	-0.010	± 13.5	± 2.9E-01 in					
0 a 69.73	69.73104	-0.013	± 14.1	± 2.9E-01 in					
0 a 70.19	70.19091	-0.011	± 14.2	± 2.9E-01 in					
0 a 74.21	74.21063	-0.008	± 15	± 2.9E-01 in					
0 a 78.89	78.89056	-0.007	± 15.9	± 2.9E-01 in					
0 a 82.25	82.25052	-0.006	± 16.6	± 2.9E-01 in					
0 a 86.17	86.17091	-0.011	± 17.4	± 2.9E-01 in					
0 a 90.34	90.34109	-0.013	± 18.2	± 2.9E-01 in					
0 a 94.27	94.27108	-0.013	± 19	± 2.9E-01 in					
0 a 98.36	98.36095	-0.011	± 19.8	± 2.9E-01 in					
0 a 99.18	99.18081	-0.0097	± 20	± 2.9E-01 in					