

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.263 °C	30.6 °C	0.337 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	35.0 %HR	4.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.355 %HR	55.2 %HR	4.845 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	74.2 %HR	3.752 %HR	± 5.0 %HR	± 1.9E+00 %HR